

ALL INDIA INSTITUTE OF MEDICAL SCIENCES
(STORES SECTION)

F. NO. 10/SO(DO)/COVID-19/2020-21/St.

DATE: 09.07.2020

Subject: Status of Evaluation of bids submitted by various firms against Short Tender Enquiry No. 10/SO(DO)/COVID-19/2020-21/St. for the purchase of VTM Kits reg.


With reference to bids for the purchase of VTM Kits submitted by various bidders against Short Tender Enquiry Vide No. 10/SO(DO)/COVID-19/2020-21/St. The Selection/rejection remarks of the departmental evaluation committee are attached for information to all concerned.



Sr. Stores Officer

(Stores Section DO)

Encl: As above.

*Prof. In-charge (CF) may kindly see for orders for
uploading on AIIMS website for information of all bidders.
Put up for orders pl.*

Prof. In-charge (CF) 



Sh. Sanjeev Kr.

10/07

COMPARATIVE STATEMENT
Specification for Viral Transport Media

Sr No.	Tender Specifications	1 International Scientific Instrument SAR Celllabs Pvt. LTD	2 Tulip Vinay Brothers	3 Genetix	4 Ginger	5 A.N. Pharma SANLI VTM	6 Chemco HI Media	7. S.K enterprises
1	Provision of two flocculated nylon swabs with break point (One Nasopharyngeal and one Throat swab).	(Not as per specifications) break point of swab is hard to break which can lead to spillage	(Not as per specifications) Have not mentioned about type of swab, Hard to break at break point	(Not as per specifications) as nature of swab is either Rayon/Dacron swab	(Not as per specifications) but length of swab after breakpoint is more than length of VTM tube which leads to non closure of tube	(Not as per specifications) Swabs are without breakpoint	(as per specifications)	(as per specifications)
2	Self standing, screw-capped tube with capacity of 10-15 ml.	(as per specifications)	(as per specifications)	(as per specifications)	(Not as per specifications) (short tube, 10 ml)	(Not as per specifications) (No screw cap)	(as per specifications)	(Not as per specifications) (10 ml)
3	At least 3 ml Hank's Balanced Salt Solution.	(as per specifications)	(as per specifications)	(as per specifications)	(as per specifications)	(as per specifications)	(as per specifications)	(as per specifications)
4	Viral Transport Medium should be clear solution with pH 7.3 ± 0.3 .	(as per specifications)	(as per specifications)	(Not as per specifications) (No Indicator)	(as per specifications)	(as per specifications)	(as per specifications)	(as per specifications)
5	Shelf life of the product should be	(as per specifications)	(as per specifications)	(as per specifications)	(as per specifications)	(as per specifications)	(as per specifications)	(as per specifications)





Dr. SIMAL KUMAR DAS
Professor
Department of Microbiology
All India Institute of Medical Sciences
New Delhi-110029

Dr. RAMA CHAUDHARY
Professor & Head
Department of Microbiology
All India Institute of Medical Sciences
New Delhi-110029

COMPARATIVE STATEMENT
Specification for Viral Transport Media

	more than or at least 12 months.							
6	Medium should also contain Cryo-protectant, to help preserve the virus.	as per specifications	as per specifications	as per specifications	as per specifications	as per specifications	as per specifications	as per specifications
7	Should contain antibiotic to prevent microbial contamination.	as per specifications	as per specifications	as per specifications	as per specifications	as per specifications	as per specifications	(Not as per specifications)
8	Should be CE/BIS marked.	as per specifications	as per specifications	as per specifications	as per specifications	as per specifications	as per specifications	as per specifications

m/s Chemco, N. Delhi was considered. It so (DO) may pl. but for Na pl.

L

Pr

Wan

Y

Re