

**DEPARTMENT OF NEUROLOGY
ALL INDIA INSTITUTE OF MEDICAL SCIENCES
ANSARI NAGAR, NEW DELHI-110029**

References: Short Rate Enquiry No.44/Neuro/KP/14-15/N-1481

Subject: Short Rate Enquiry for Horizontal Gel Electrophoresis with Power Pack under Department of Neurology, AIIMS, New Delhi.

A Short rate enquiry is hereby floated/invited for purchase of the following **Horizontal Gel Electrophoresis with Power Pack** in Department of Neurology, AIIMS, New Delhi.

Kindly supply the rates of following items with full terms and conditions.

S.No	Specifications for Horizontal Gel Electrophoresis
1.	A horizontal electrophoresis system with gel caster should be able to accommodate the gel tray of the sizes of 15x10cm or 15x7cm and should come with safety lid.
2	The system should include UV transparent gel tray of size 15x10cm with integrated fluorescent ruler.
3	The system should include tape free gel casting module for leak free operations.
4	The system should include casting gates to allow hassle free casting of gels directly in the electrophoresis cells.
5	The system should include four sets of 1.5mm 15- & 20-well fixed height combs each.
6	The sample throughput of the system should be 1-60 samples (2 x 30 well).
7	A system should have the option for adjustable height combs with comb holders.
8.	The base buffer volume should be as low as 650ml
9.	Migration rate of Bromophenol Blue dye should be similar to 4.5cm/hr(at 75 V).
10.	A system should have a lock-in design to run precast agarose gels.

S. No	Specifications for Power Pack
1.	High Quality power supply that offers simple programming .Intended for Horizontal gel Electrophoresis.
2.	Output Specifications: Volts - 10-300V, Current - 4-400mA, Power - 75 W
3.	Type of output -constant voltage, constant current, constant power with automatic crossover.
4.	Number of output Jacks: upto 4 identical electrophoresis cells may be connected in parallel to power supply.
5.	Timer control 1 Min -900 mins
6.	Ability to Pause & Resume function should be available
7.	Option of automatic recovery should be present
8.	System should be Stackable case with adjustable viewing angle via flipdown legs.
9.	Automatic detection for NO load, sudden load change, short circuit, over voltage conditions.
10.	3 Digit LED display should be available.
11.	Operating Conditions should be suitable for Indian temperature conditions

The Quotations should be addressed to Prof. Kameshwar Prasad, Prof. & Head, Department of Neurology, Room No:- 91, Clinical Epidemiology Unit, Near Examination section, AIIMS, New Delhi-110029 and reach on or before **Saturday, 21/03/2015 – 12:00 noon** by post/courier. No quote shall be accepted by hand/in person.

Last Date of submitting of quotation : Saturday, 21/03/2015 – 12:00 noon

Date of Opening of quotation : Monday, 23/03/2015

TERMS AND CONDITIONS:

1. Only authorized distributors should submit the quote and recent authorization letter must be provided.
2. The total/ combined prices of all the items will be considered as the final price quote for price comparison purposes.
3. The firm should provide the list of government institutions where they have supplied the similar items.
4. The quantity can be increased or decreased as per requirement.
5. The firm must be registered and having TIN no./Registration no. Please mentioned TIN/ registration no. on the bid or else the quotation will be treated as cancelled.
6. The make/model of the article offered should invariably be quoted. Quotation should be typed/ written in ink. No overwriting or erased entries should be there in the quotation. The rate should be valid at least one Month from the date of opening of quotations.
7. VAT/ Sales tax or any other kind of tax(s) must be mentioned separately as applicable on the item. In case no sales tax is chargeable, prices must be quoted as NET price.
8. The under signed has reserves the right to call the firm for physical demonstration of item to verify the quality of the offered item. The firm must have to demonstrate the product within 2 days of requisition, failing which the bid will be rejected.
9. The firm should be competent to supply the item at the mentioned place/ location within seven days after issuing the Supply Order
10. The payment will be made electronically viz RTGS/NEFT after delivery of the item as per rules of the AIIMS Research Section in the name of the firm only. The following information should be also mentioned in the invoice:
 - a)Name of the beneficiary:
 - b)Type of Account :
 - c)Account No. of the beneficiary:
 - d)IFSC code of the bank/ branch:
11. At any time, any part of the rate enquiry which generally does not fulfill the conditions stipulated in the rate enquiry can be cancelled/ rejected.