

**OFFICE OF THE DIRECTOR
ALL INDIA INSTITUTE OF MEDICAL SCIENCES
ANSARI NAGAR, NEW DELHI-29
(DEPARTMENT OF PHARMACOLGY)**

R.E.No.9/Pharma/14-15

Dated 10/02/2015

Sealed quotations are invited by post/per bearer i.e. (A) Technical bid (B) Price Bid. Both bids should be sealed in separate envelopes clearly mentioning on the top of envelope. Keep both bids in envelope sealed, for the supply of items detailed below which are required for Department of Pharmacology, AIIMS, New Delhi-110029.

The quotations should reach this office on or before 25-02-2015 at 11.00 AM duly sealed & marked Tender/R.E. No. 09/**Pharma/14-15**. All quotations should be printed. Over writing and erased entries will not be considered and treated as deleted entries from the quotations. The validity of the quotation should be minimum for three months.

The tenders are to be deposited with bid security of Rs. 5,000/= (Refundable) (Rupees Five thousand only) through a Demand Draft/Pay order drawn in favour of the "Director, AIIMS, New Delhi", along with the quotation failing which the quotation will not be considered. In case of failure to comply with the supply order in terms of the contract, the order will be placed to the next successful bidder and the bid security money deposited will be forfeited. The approved firm has to supply the item/material within a period of 15 days after issue of confirmed supply order.

The firm must quote with make, Model and the catalogue. To be noted that without this, the quote will become invalid.

The quotations will be opened on 26-02-2015 at 11.30 AM in the Room No 4013, Department of Pharmacology, AIIMS, New Delhi- 110029 in the presence of the tenderers, whosoever are present.

Sr. No. Nomenclature

Qty. Req.

Refrigerator 400 lit Double Door with minimum (4 Star) Rating

6 nos

Specification for Refrigerator

1. Capacity should be minimum 380 litres.
2. Should have Double Door.
3. Should have Locking System, Inbuilt stabilizer, minimum 4 star ratings for energy saving.
4. Must provide the make, Model and catalogue.
5. ISO/CE certificate
6. Warranty should be 2 Years.

Dr. D.S.Arya
Prof. & O/I (Store)

Dr. Y.K.Gupta
Prof. & Head

Note:- Sample of the quoted item must be submitted along with quotation failing which offer will not be considered for acceptance.