

DEPARTMENT OF REPRODUCTIVE BIOLOGY
All India Institute Of Medical Sciences
Ansari Nagar, New Delhi - 110029

Rate Enquiry No. 01/CRIA/R.B./M&E(Plan)/RE/2016-17

Dated: 14/09/2016

Quotations are invited by post / courier / hand for the supply of items mentioned below which are required for Department of Reproductive Biology, AIIMS.

Rotary Vacuum Evaporator – Quantity: 1

* The detailed technical specifications for the item are given in separate page below.

The quotations should be submitted to Professor & Head, Department of Reproductive Biology, Room no. 2083, IInd floor, Teaching Block, AIIMS, New Delhi latest **by 30.09.2016 upto 5.00 PM**. All quotations should be typewritten or in ink. All over writing and erased entries will be deleted from the quotations.

Terms & Conditions:

- The quotation should be duly sealed in an envelope marked “R.E. No. 01/CRIA/R.B./M&E(Plan)/RE/2016-17 for Rotary Vacuum Evaporator”.
- The technical bid and price bid should be submitted in separate envelopes with the envelopes clearly marked on top as “technical bid” or “price bid”. If bids are not submitted in this format, they will be rejected.
- Price bid of only those firms will be opened whose technical bid is approved by the tender committee.
- The firm should submit the following along with the quotation:
 1. Authority letter from manufacturer/principal for supplying item without which tender will not be considered.
 2. Price Reasonable Certificate.
 3. An undertaking that the firm has not been blacklisted in the past by any Institution, Government/Private.
 4. The names & contact numbers of AIIMS/Govt. departments where the supplier/dealer has supplied the item (user list).
 5. The make of the article should be clearly mentioned and relevant literature with picture/ original catalogue should be attached.
 6. Information for electronic payment viz RTGS / NEFT – Name of the beneficiary; Account number of the beneficiary; IFSC code of the bank / branch.
- Rates should be valid for at least THREE MONTHS.
- If applicable, VAT / Sales Tax or any other kind of tax must be mentioned separately against each item.
- If required, the firm should be able to give demonstration of the item within one week of requisition.
- The firm must be able to install the equipment on the site, free of cost.
- The firm must supply the items within the date mentioned on the supply order copy.
- Payment will be made after receiving the items in good condition & satisfactory installation.

Rotary Vacuum Evaporator – Quantity: 1

- Microprocessor controlled motor to provide smooth variable speed.
- It should be capable of adjustable rotation speed of 20 to 250 rpm.
- Equipped with an action jack lift for manual lifting of the evaporating flask.
- Digital display for rotation speed.
- It should have cold trap condenser with all accessories to be included.
- Maximum flask size range 50 to 4000 mL.
- Slandered joints 29/32, 24/40 – 2 each and flask 100ml, 250ml, 1L, 2L and 4L should be provided.
- Heating Bath.
- Integrated in Heating Bath should allow an easy positioning of bath along horizontal axe for individual.
- Digital temperature controller and display allows adjusting the set temperature and indicates actual temperature.
- Protection class IP21 Controlled temperature range 20 – 95 °C with 1 °C increment.
- Temperature regulation accuracy $\pm 2^{\circ}\text{C}$.
- Over-temperature protection.
- Vacuum pump: Appropriate Integrated vacuum pump.
- Recalculating chiller: Integrated water chiller for condenser for cold trap.
- Operation voltage 220 – 240 V.
- International certification/ CE/ FDA.
- Appropriate UPS should be provided with 30 min or more backup.
- Five year comprehensive Warranty and 5 year AMC charges after warranty year wise.
- User list and certificate of satisfactory service.