

DEPARTMENT OF REPRODUCTIVE BIOLOGY
All India Institute Of Medical Sciences
Ansari Nagar, New Delhi - 110029

Rate Enquiry No. 06/CRIA/R.B./M&E(Plan)/RE/2016-17

Dated: 20/01/2017

Quotations are invited by post / courier / hand for the supply of items mentioned below which are required for Department of Reproductive Biology, AIIMS.

Serum Separation Centrifuge – Quantity: 1

* The detailed technical specifications for the item are given in separate page below.

The quotations should be submitted to Professor & Head, Department of Reproductive Biology, Room no. 2083, IInd floor, Teaching Block, AIIMS, New Delhi latest **by 04.02.2017 upto 1.00 PM**. All quotations should be typewritten or in ink. All over writing and erased entries will be deleted from the quotations.

Terms & Conditions:

- The quotation should be duly sealed in an envelope marked “R.E. No. 06/CRIA/R.B./M&E(Plan)/RE/2016-17 for Serum Separation Centrifuge”.
- The technical bid and price bid should be submitted in separate envelopes with the envelopes clearly marked on top as “technical bid” or “price bid”. If bids are not submitted in this format, they will be rejected.
- Price bid of only those firms will be opened whose technical bid is approved by the tender committee.
- The firm should submit the following along with the quotation:
 1. Authority letter from manufacturer/principal for supplying item without which tender will not be considered.
 2. Price Reasonable Certificate.
 3. An undertaking that the firm has not been blacklisted in the past by any Institution, Government/Private.
 4. The names & contact numbers of AIIMS/Govt. departments where the supplier/dealer has supplied the item (user list).
 5. The make of the article should be clearly mentioned and relevant literature with picture/ original catalogue should be attached.
 6. Information for electronic payment viz RTGS / NEFT – Name of the beneficiary; Account number of the beneficiary; IFSC code of the bank / branch.
- Rates should be valid for at least THREE MONTHS.
- If applicable, VAT / Sales Tax or any other kind of tax must be mentioned separately against each item.
- If required, the firm should be able to give demonstration of the item within one week of requisition.
- The firm must be able to install the equipment on the site, free of cost.
- The firm must supply the items within the date mentioned on the supply order copy.
- Payment will be made after receiving the items in good condition & satisfactory installation.

Serum Separation Centrifuge – Quantity: 1

- Non Refrigerated Bench Top Centrifuge for serum separation.
- Microprocessor controlled table top non–refrigerated centrifuge with LCD display of RPM, RCF, temperature, time counter, Rotor model & Acceleration/deceleration curve simultaneously.
- Speed at least 4000 – 5000 RPM.
- Can be set in both RPM and RCF.
- Runtime 1 – 60 min and / or continuous run with timer.
- Maintenance free brushless induction motor drive.
- Automatic rotor recognition and motorized locking of lid with over-speed protection.
- RPM adjustment in increment/decrement of 100 RPM or better.
- Smart knob to easily adjust RPM/RCF for routine work.
- Quick spin key for short operation.
- At least 3 programme memory and one current programme.
- Real time modification of parameter during centrifugation.
- Swing out rotor for 13x75/100mm (5-6 ml) vaccutainer or tube for at least 120 tubes in one run (preference will be given to those with higher number of tubes).
- Active imbalance identification and cut off.
- Manufactured according to international safety regulation.
- The Noise should be less than 60dB at maximum speed.
- The product should be CE certified/or international certification.
- Five year comprehensive Warranty and 5 year AMC charges after warranty year wise.
- Appropriate voltage stabiliser with quality certification.
- User list and certificate of satisfactory service.