

**CENTRE FOR COMMUNITY MEDICINE
ALL INDIA INSTITUTE OF MEDICAL SCIENCES
ANSARI NAGAR, NEW DELHI-110029**

Ref No. RE:1339/Oct/Enq06

Specification for Swabs for collection of nasal and throat sample

1. It should have polyester fiber-tipped applicator.
2. Polyester tip fibers should provide efficient absorption and complete release of liquid samples for testing of DNA & RNA.
3. Swab shaft should be plastic material of diagnostic grade of length size: 6 inches.
4. It should be ready to use sterile, DNase and RNase free.
5. Packing : two swabs in one peel bag.



डॉ. आनंद कृष्णन / Dr. Anand Krishnan
आचार्य Professor
सामुदायिक चिकित्सा विभाग केन्द्र
Centre for Community Medicine
अ.भा.आ.सं., नई दिल्ली-११००२९
A.I.M.S., New Delhi-110029