

**OBSTETRICS & GYNAECOLOGY**  
**AIIMS, ANSARI NAGAR, NEW DELHI-29**

Ref. 1/O & G/2017-18

Dated 28/08/2017

Subject:- Purchase of Cervix-Coagulator Machine and Therapy Probe for the Department of Department of Obst. & Gynae, AIIMS, New Delhi-29 on Proprietary Basis **Inviting Comment thereon.**

The department of Obst & Gynae wants to purchase of Cervix-Coagulator Machine and Therapy Probe each one No. from M/s. Wisap Medical Technology GmbH, Germany, on proprietary basis. The proposal submitted by the local authorized dealer M/s. Kriticare Solutions, Delhi.

The above documents are being uploaded for open information to submit objections, comments, if any from any manufacture regarding proprietary nature of the Cervix-Coagulator Machine and Therapy Probe within 20 days from the date issue/uploading the notification giving **reference No. 1/O & G/2017-18.**

The comments should be sent to Professor & Head, department of Obst. & Gynae, 3<sup>rd</sup> Floor, Teaching Block, AIIMS, New Delhi on or before **20/09/2017 upto 05.00 P.M.** failing which it will be presumed that any other vendor is having no comment to offer and case will be decided on merits.

(PROF. ALKA KRIPLANI)  
PROF. & HEAD,  
DEPTT. OF OBST. & GYNAE

Encl:-

Quotation, Lowest Price Certificate  
Technical Specification for Cervix Coagulator  
PAC Certificate  
Letter of authorization

*Neer*  
*28*  
Dr. Neerja Bhatla  
MD, MRCOG  
Professor  
Deptt. of Obst. & Gynae.  
AIIMS, New Delhi-29  
**DR. NEERJA BHATLA**  
Professor  
Dept. of Obst. & Gynae.  
A.I.I.M.S., New Delhi



## KRITICARE SOLUTIONS

403, SANTOSH APARTMENTS, PLOT NO.39B,  
SECTOR-6, DWARKA DELHI - 110075  
E-mail:ritu\_barreja@yahoo.com  
Phone :9958366200

TIN : 07296979382

### Quotation for Cervix Coagulator

Item Code	Description	Unit Price(INR)
6001	Cervix-Coagulator(including Instrument cable)	3,98,505.80
6002	Therapy Probe	88,759.13
6003	Therapy Probe	88,759.13
6004	Therapy Probe	88,759.13
6005	Therapy Probe	88,759.13
6006	Therapy Probe	94,560.38
6007	Therapy Probe	88,759.13
6008	Therapy Probe	88,759.13
6009	Therapy Probe	99,549.45

This is the lowest quoted rate with two years of warranty of the coagulator.

#### Terms & Conditions:

1. GST@12% extra or as applicable
2. Payment terms: 100% payment on delivery.& successful installation.
3. Delivery: 8-10 weeks from the date of PO.
4. Warranty: Two years from the date of installation for the coagulator(6001)
5. Warranty not applicable on probes or accessories (instrument cable, Power cord) or consumables.
6. Physical damages & mishandling of equipment, or damage due to voltage fluctuation not covered in warranty.
7. Installation & training will be carried out once by factory trained Engineer at NO EXTRA CHARGE.
8. CMC CHARGES: 5% of the PO value + taxes as applicable (including labour, Excluding probes, consumables& accessories). 1% YoY escalation in base value of the CMC.

*Ritu*  
**KRITI CARE SOLUTIONS**  
403, SANTOSH APARTMENTS  
PLOT NO.39B, SECTOR-6,  
DWARKA, DELHI-110075  
TIN No.: 07296979382



*Neerja*  
**Dr. Neerja Malhotra**  
MD, MRCOG  
Professor  
Deptt. of Obst. & Gynae.  
AIIMS, New Delhi

*Neerja*  
**DR. NEERJA BHATLA**  
Professor  
Dept. of Obst. & Gynae.  
A.I.I.M.S., New Delhi

**WISAP**

medical technology

To Whom It May Concern

MANUFACTURER'S AUTHORIZATION LETTER

Brunnthal/Hofolding, 22.08.2017

This is to confirm that **KRITICARE SOLUTIONS**, having its registered office at 403, Santosh Apartments, sec 6, plot No. 39B, Dwarka, Delhi-110075, INDIA is authorized to quote and supply the Cervix Coagulator System (= Cold Coagulator System) (ref. 6001 + accessories 6020, 6002, 6003, 6004, 6005, 6006, 6007, 6008, 6009, 6010) from WISAP Medical Technologies having its principal office at Fichtenstrasse 27, 85649 Brunnthal/Hofolding, GERMANY.

This letter of authorization is effective on the date of issue and shall be valid till 31.12.2018.

Yours faithfully,

For and on behalf of

WISAP Medical Technology GmbH

**WISAP®**

WISAP Medical Technology GmbH  
Fichtenstrasse 27 - 85649 Brunnthal - Hofolding  
Tel.: +49 (0)8104 / 8908-0 ; Fax: +49 (0)8104 / 8908-90  
info@wisap.de ; www.wisap.de

Sandra Seywald-Schmidt  
Head of Sales

*Neer*  
Dr. Neerja Bhatla  
MD, MRCOG  
Professor  
Dept. of Obst. & Gynae.  
AIIMS, New Delhi-29

*se*  
DR. NEERJA BHATLA  
Professor  
Dept. of Obst. & Gynae.  
A.I.I.M.S., New Delhi

WISAP Medical Technology GmbH  
Fichtenstraße 27  
85649 Brunnthal  
Germany  
Phone: +49(0)8104/8908 - 10  
Fax: +49(0)8104/8908 - 90

Sitz der Gesellschaft  
Brunnthal OT Hofolding  
Amtsgericht München  
HRB 191426  
Geschäftsführer:  
Stephan Hilgers

Bankverbindung  
Commerzbank München  
Kto.: 662914100  
BLZ 70040041  
IBAN DE65700400410862914100  
BIC/SWIFT COBADEF700

## PROPRIETARY CERTIFICATE

WE HEREBY CERTIFY THAT THE FOLLOWING EQUIPMENT (EQUIPMENT NAME = COLD COAGULATOR OR CERVIX COAGULATOR, SEE PRODUCT IMAGE) IS EXCLUSIVELY MANUFACTURED BY US. WE ALSO CERTIFY THAT THE BELOW MENTIONED PRODUCT INCLUDING ITS SPARE PARTS AND ACCESSORIES IS NOT BEING MANUFACTURED BY ANY OTHER MANUFACTURER IN THE WORLD AND ITS QUALITY IS BEING CERTIFIED BY CE.

### EQUIPMENT NAME

COLD COAGULATOR OR CERVIX COAGULATOR

### REFERENCE NO.

6001

### IMPORTANT KEY FEATURES

- REUSABLE THERAPY PROBES
- PORTABLE UNIT
- FAST TREATMENT TIME
- HIGH EFFICACY

YOURS FAITHFULLY

For WISAP MEDICAL TECHNOLOGY GmbH



The certificate is valid until 31<sup>st</sup> of December 2017.

**WISAP®**

WISAP Medical Technology GmbH  
Fichtenstrasse 27 - 85649 Brunnthal - Hofolding  
Tel.: +49 (0) 8104 / 8908-0 Fax: +49 (0) 8104 / 8908-90  
[info@wisap.de](mailto:info@wisap.de) [www.wisap.com](http://www.wisap.com)

*Lei Zhang*

Lei Zhang, 16.02.2017

(Head of Regulatory / Quality Department)

*Dr. Neerja Bhatla*  
MD, MRCOG  
Professor  
Dept. of Obst. & Gynaecology  
AIIMS, New Delhi-29

DR. NEERJA BHATLA  
Professor  
Dept. of Obst. & Gynaecology  
A.I.I.M.S., New Delhi

WISAP Medical Technology GmbH  
Fichtenstrasse 27  
85649 Brunnthal  
Germany  
Phone: +49(0) 8104/8908 - 10  
Fax: +49(0) 8104/8908 - 90

Sitz der Gesellschaft  
Brunnthal OT Hofolding  
Amtsgericht: München  
HRB 191426  
Geschäftsführer:  
Stephan Hilgers

Bankverbindung  
Commerzbank München  
Kto: 662914100  
BLZ: 70040041  
IBAN: DE65700400410662914100  
BIC/SWIFT: COBADEFF700



## KRITICARE SOLUTIONS

403, SANTOSH APARTMENTS, PLOT NO.39B,  
SECTOR-6, DWARKA DELHI - 110075  
E-mail:ritu\_barreja@yahoo.com  
Phone :9958366200

TIN : 07296979382

### Technical Specification for Cervix coagulator/Thermo Coagulator

#### Details of the Equipment:

- The cervix coagulator should meet following requirements
- Cervix-Coagulator/Thermo coagulator/Cold coagulator incl. country- specific power cord along with Instrument Cable, with two coupling plugs, silicone, L = 1.5 m
- Temperature desired should be achieved in 30 sec & takes approx. 2 minutes to cool.
- Treatment time should be less than a minute.

#### ELECTRICAL SPEC.

Mains connection	Standard: 230V - 50 /60 Hz, need continuous supply of electricity.
Power input	66 VA
Range of temperature	60-120°C
Regulation of temperature	Step less, including temperature control

#### MECHANICAL SPEC.

Dimensions	w x h x d = 270 x 120 x 190 mm
Casing	Stainless steel, aluminium / white coated
Weight	3-4 kg
Thermo-probes	Stainless steel, PTF Coated for anti-adhesion-effect, FDA approved


*RF*  
KRITICARE SOLUTIONS  
403, SANTOSH APARTMENTS  
PLOT NO. 39B, SECTOR-6  
DWARKA, DELHI-110075  
TIN No.: 07296979382

*Neer*  
*Re*  
Dr. Neena Bhatla  
MD, MRCOG  
Professor  
Dept. of Obst. & Gynaec.  
AIIMS, New Delhi-28  
DR. NEERJA BHATLA  
Professor  
Dept. of Obst. & Gynaec.  
A.I.I.M.S., New Delhi



SYSTEM-FEATURES	
Anti-Stick-coating therapy probes	Teflon (FDA approved)
"Cold-Coagulation" of glandular tissue	Mild and efficacious method of destroying cervical lesions, in which no current is flowing through the patient but the cell is destroyed by temperature dissemination in to the tissue. This leads to a lower incident rate after treatment, but offers same efficacy rate.as other treatments
Range of temperature of therapy probes:	60-120°C
Thermo-probes should be able to treat	Large area changes of the cervix (70-100°C)
	Dehiscent cervical os (70-100°C)
	Narrow cervical os (70-100°C)
	Endocervical- / endouterine therapy (60-70°C)
	Endocervical- / endouterine therapy (60-70°C)
	Large area changes of the cervix (70-100°C)
	Dehiscent cervical os (70-100°C)
	Haemostasis after conisation (LEEP) (120°C)
	Large area changes of the cervix (100°C)
CERTIFICATION	
Approvals	CE labelled.
Safety Spec.	Classified according to 93/42/EWG IIa, Protection class I, Type BF
STERILIZATION METHOD	
Sterilization control unit	Wipe-disinfected
Sterilization of therapy probes	Autoclavable (134°C)

*Ritu*  
**KRITI CARE SOLUTIONS**  
 403, SANTOSH APARTMENTS  
 PLOT NO.39B, SECTOR-6  
 DWARKA, DELHI-110075  
 TIN No.: 07296979382

*Neer*  
 Dr. Neerja Bhatla  
 MD, MRCOG  
 Professor  
 Deptt. of Obst. & Gynae.  
 AIIMS, New Delhi-29

*Ne*  
**DR. NEERJA BHATLA**  
 Professor  
 Dept. of Obst. & Gynae.  
 A.I.I.M.S., New Delhi