


Rate Enquiry No - 3/Haem/15-16


DEPARTMENT OF HEMATOLOGY, A.I.I.M.S.

SPECIFICATION FOR THERMOSTAT HOT PLATE

Quantity : 1 no.

- Should have a ambient ^{°C} +5 UP to 200°C Stability +0.5°C
- Should have a Electronic digital temperature and time control display
- Should have suitable for thermal tests (not suitable for boiling tests)
- Should have surface plate coated with P.T.F.E. for protection against chemicals
- Should have a indicator alarm if the plate is over temperature
- Should have ISO Certificate.




डॉ. सीमा त्यागी/Dr. Seema Tyagi
आचार्य/Professor
रुधिर विज्ञान विभाग/Department of Hematology
अ.भा.आ.सं. नई दिल्ली 20/A IIMS New Delhi-29

Achary HOD