



COMPREHENSIVE RURAL HEALTH SERVICES PROJECT
CENTRE FOR COMMUNITY MEDICINE
ALL INDIA INSTITUTE OF MEDICAL SCIENCES
BALLABGARH - 121004 (HR) / Ph. 0129-2241362/Fax. 0129-2211227

S.R.E.No.10/17-18/CRHSP-AIIMS /Ballabgarh

Dated the: 02.11.2017

For Purchase of : ENT MICROMOTOR DRILL SYSTEM WITH HANDPIECE & BUR - Qty - 02 Nos.

The sealed quotations are invited by post/per bearer (under two bid system) in two separate sealed covers duly marked as "PRICE BID & TECHNICAL BID" (as the case may be) and both packed in other single envelope mentioning name of the quoted item for the supply of item detailed below which are required for CRHS Project, Ballabgarh.

The quotations duly signed and marked as "SRE No.10/CRHSP/Ballabgarh/17-18" (PRICE BID & TECHNICAL BID separately) should reach the office of the Professor In-charge, CRHSP, Ballabgarh on or before 17.11.2017 at 5.00 P.M. All quotations should be type written or written with indelible ink. All over writing and erased entries will not be considered and treated as deleted entries from the quotations. The rates should be valid at least for one year. A clause by clause compliance statement should be submitted, failing which the quotation will not be considered.

The tenders are required to enclose with the quotation an Earnest Money of Rs.5,000/- (Rupees Five Thousand only) through a demand draft/pay order drawn in favor of the "Director, AIIMS", failing which the quotation will not be considered. In case of failure to comply with the supply order in terms of the contract, purchase will be made at the cost and risk of the contractors and or liquidated damages may be recovered from them. The period within which the goods can be supplied from the date of issue of the supply order should be specified.

The offers submitted as quality A, B, C, & D etc., will be summarily rejected, any alteration/over writing in the specification / rates shall not be considered at all.


The opening of quotations/tenders (technical bid only) will be intimated to the bidders separately. The quantity shown in the Short Rate Enquiry can be increased or decreased to any extent depending upon the actual requirement.

Terms and Conditions:-

1. Machine should be CE/USA FDA approved.
2. Two years comprehensive warranty & 03 Yers CMC.
3. Standard AIIMS terms & conditions will also apply.

The firm should submit the following:-

1. A clause by clause compliance statement should be submitted.
2. Price Reasonable Certificate.
3. An undertaking on the official letterhead of the company that the firm has not been black listed.
4. An undertaking on the official letter head of the company that no Vigilance case is pending against the firm.


(DR. SHASHI KANT)
PROFESSOR & HEAD,
CCM & Prof-In-Charge, CRHS
PROJECT, BALLAGARH

Technical Specifications ENT Micromotor Drill System With Handpiece & Burrs

Following are the minimum requirements, Products offered must these parameters herein

Sr	Technical specifications/composition of tender enquiry	Compliance on each parameter with detailed substantiation how the offered product meets the requirements. (Simply writing a YES/Complied As per BIS/CE/ISO terms is not allowed)	Remarks
1,	ENT Micro drilling Systems		
	a. High speed not less than 60000 rpm		
	b. Dual speed control from control box as Well as foot pedal		
	c. Both modes forward and reverse cutting		
	d. Inbuilt pump for irrigation with minimum 5cc to 30cc		
	e. Flexible irrigation setting for international & Continuous flow		
2.	Micro Motor Hand Pieces		
	a. Rated for Minimum 60000 rpm		
	b. Ball bearing type as to generate minimum heat & vibration		
	c. With attached irrigation pipes.		
	d. Standard coupling for micrometer		
	e. Straight hand pieces – 02 Nos.		
	f. Contra angled handpiece – 02 Nos.		
3.	Burs		
	a. Tungsten carbide burrs cutting set of 12 varying from 0.6mm to 7 mm standard length of 70 mm. Round, conical and cylindrical type		
	b. Diamond burrs complete set of six varying from 0.6mm to 7 mm length 70mm-round, conical and cylindrical type.		
	c. Oil spray for hand piece with nozzle		
	d. Two years (02) warranty and good after sales service.		
	e. Should be CE/US, FDA approved		