

DEPARTMENT OF MICROBIOLOGY  
ALL INDIA INSTITUTE OF MEDICAL SCIENCES

Rate Enquiry No. 3 /Micro/2016-2017

Dated 29 -10-2016

Quotations are invited for the supply of items mentioned below which are required for Microbiology Department AIIMS.

1. Shaker Incubator – 1 No.
2. Binocular Compound Microscope – 2 Nos.
3. Water Bath – 1 No.
4. Weighing Balance – 1 No.
5. Horizontal Gel Electrophoresis unit – 1 No.
6. Electrophoresis Power Pack – 1 No.
7. pH Meter – 1 No.
8. LED Fluorescent Microscope – 1 No.
9. Electronic Balance – 1 No.
10. Autoclave – 1 No.

The quotations should be submitted to Prof. & Head Department of Microbiology Department Room No.2060 IInd Floor, Teaching Block, AIIMS, New Delhi by 19-11-2016 upto 4.00 pm. The quotations will be opened on 21-11-2016 at 3.00 pm Room No.2093. All quotations should be typewriter or in ink. All over writing and erased entries will be deleted from the quotations. The quotation should be duly sealed in an envelope marked R.E.No. 3 /Micro/2016-2017.

Terms & Condition:

1. The quotation for each item should be submitted in a separate with the name of the item clearly written on top of the envelope.
2. The Rate Enquiry must quote for Five Years Comprehensive Warranty.
3. Authority letter from manufacturers/principal for supplying item.
4. Price reasonable certificate.
5. An undertaking that the firm has been blacklisted in the past by any Institution, Government/Private.
6. The names & contact numbers of AIIMS/Govt. Departments where the supplier/dealer has supplied the equipment (user list).
7. Information for electronic payment viz RTGS/NEFT – Name of the beneficiary, Account number of the beneficiary, IFSC code of the bank/branch.
8. If applicable, VAT/sales tax or any other kind of tax must be mentioned separately against each item.
9. The firm must be able to install the equipment on the site, free of cost.
10. The firm must supply the items within the date mentioned on the supply order copy.



Dr. Arti Kapil  
Professor of Microbiology  
Officer Incharge Store



Dr. Gita Satpathy  
Prof. & Head  
Department of Microbiology

### Shaker incubator

1. Should be provided with Temperature control range from 5°C above ambient to 60°C control accuracy  $\pm 0.1^\circ\text{C}$ .
2. Should be able to maintain 37°C or below during summer months when ambient temperature is 40-45°C.
3. Should be provided with shaking Speed: 20-800 rpm with control accuracy of  $\pm 2-4\text{rpm}$ .
4. Should be provided with timer of up to 99.9 Hrs.
5. Should be provided with approximate dimensions (WxDxH): 20-30 x 20-30 x 10-25cm and weight of the basic device  $\leq 10-15\text{ Kg}$ .
6. Should be provided with Power supply: 100-240V, 50-60 Hz.
7. Should be provided with Large easy to read display screen clearly indicates speed, running time, alarm condition and temperature character and Display is clearly visible from all angles across the room, and in the dark.
8. Should be provided with Safety Features i.e. there should be over temperature protection. The system should have facility to calibrate the temperature and speed via keypad.
9. Should be provided with electronic and mechanical components should be enclosed, protected from accidental spills.
10. Should be provided with automatic restart after power interruption with non volatile memory.
11. Should be provided with It should come with Universal Platform having capacity to holds assortment of various size of flask sizes upto 2-5 Ltrs and test tube racks. System should be supplied with 250 ml clamps (5 Nos.), 500 ml clamps (05 Nos.), 1000 ml (01-03 Nos.) and 2000 ml (01-03Nos)
12. Should be provided with comprehensive warranty of 5 years.
13. Should be ISO 9001 and 14001 certified. The instrument should be an original make from the company and not assembled locally.

Certified that specifications are broad based, general in respect to the requirement and not suit to any particular firm / brand.

Refused by  
Tamilogay

डॉ. तनु सगर  
Dr. TANU SAGAR, MD  
सहायक अध्यापक/Assistant Professor  
सुखन जेय विज्ञान विभाग/Deptt. of Microbiology  
अखिल भारतीय आयुर्विज्ञान संस्थान/A.I.I.M.S.  
अंसारी नगर, नई दिल्ली/Ansan Nagar, New Delhi-29

Checked by

Dr. RAMA CHANDRAN  
Professor  
Department of Microbiology  
All India Institute of Medical Sciences  
Ansan Nagar

Accepted