

DEPARTMENT OF MICROBIOLOGY
ALL INDIA INSTITUTE OF MEDICAL SCIENCES

Rate Enquiry No. 3 /Micro/2016-2017

Dated 29 -10-2016


Quotations are invited for the supply of items mentioned below which are required for Microbiology Department AIIMS.


1. Shaker Incubator – 1 No.
2. Binocular Compound Microscope – 2 Nos.
3. Water Bath – 1 No.
4. Weighing Balance – 1 No.
5. Horizontal Gel Electrophoresis unit – 1 No.
6. Electrophoresis Power Pack – 1 No.
7. pH Meter – 1 No.
8. LED Fluorescent Microscope – 1 No.
9. Electronic Balance – 1 No.
10. Autoclave – 1 No.

The quotations should be submitted to Prof. & Head Department of Microbiology Department Room No.2060 IInd Floor, Teaching Block, AIIMS, New Delhi by 19-11-2016 upto 4.00 pm. The quotations will be opened on 21-11-2016 at 3.00 pm Room No.2093. All quotations should be typewriter or in ink. All over writing and erased entries will be deleted from the quotations. The quotation should be duly sealed in an envelope marked R.E.No. 3 /Micro/2016-2017.

Terms & Condition:

1. The quotation for each item should be submitted in a separate with the name of the item clearly written on top of the envelope.
2. The Rate Enquiry must quote for Five Years Comprehensive Warranty.
3. Authority letter from manufacturers/principal for supplying item .
4. Price reasonable certificate.
5. An undertaking that the firm has been blacklisted in the past by any Institution, Government/Private.
6. The names & contact numbers of AIIMS/Govt. Departments where the supplier/dealer has supplied the equipment (user list).
7. Information for electronic payment viz RTGS/NEFT – Name of the beneficiary, Account number of the beneficiary, IFSC code of the bank/branch.
8. If applicable, VAT/sales tax or any other kind of tax must be mentioned separately against each item.
9. The firm must be able to install the equipment on the site, free of cost.
10. The firm must supply the items within the date mentioned on the supply order copy.


Dr. Arti Kapil
Professor of Microbiology
Officer Incharge Store


Dr. Gita Satpathy
Prof. & Head
Department of Microbiology

Specifications for Power Pack

1. Should have output Range (Programmable):
 - a. 5-280 V, fully adjustable in 1 V increments.
 - b. 0.01-3.0A, fully adjustable in 0.01A increments.
 - c. 1-300W, fully adjustable in 1 W increments
2. Type of output: Constant voltage, constant current or constant power with automatic crossover
3. Digital display with 16 character, 2 lines
4. Output Terminal: 4 recessed sets in parallel
5. Timer: 0-999 minutes
6. Operating conditions: 0-40 degree, 0-95% humidity in absence of condensation
7. Input protection: Fuse on hot and neutral
8. Input power (nominal): 100-120/220-240 V, AC, autoswitching
9. Dimensions: 20-25 x 25-30 x 5-10 cm (WxDxH)
10. Safety measures: No lead detection, sudden load change detection, arc detection, overload/short circuit detection, over voltage protection
11. Should provide 5 yrs of comprehensive warranty

Certified that the specifications are broad based, general in respect to the

requirement and not suit to any particular firm/brand.

Prepared by

Dr. B. K. Sharma

Dr. B. K. Sharma M.D.

Professor

Dept. of Microbiology

AIIMS, New Delhi

checked by
[Signature]
[Signature]

[Signature]