

**Specifications for One Additional Channel for MR compatible Infusion Pump
(Iradimed MR Safe Pump, Model MRIdium 3860) for Upgrade of to 2 Channel
Pump existing in the Department of NMR, AIIMS, New Delhi**

1. The upgrade should be for Iradimed MR Safe pump already in use at the institute. It should be adaptable for both uses as a syringe pump as well as a volumetric pump operable on 100 to 240 VAC +/- 10%, 50/60 Hz.
2. It should be designed for use in MR environment and not adopted for the purpose (should not be a non MR pump in an RF cage).
3. It should have a non magnetic ultrasonic motor to provide accurate fluid delivery from infants to adults.
4. The second channel should use the programming of existing main pump (with a 10 digit keypad entry system and LCD display), with dual channel programming options for simultaneous use of both the channels.
5. It should be 10000 Gauss compliant so that it can be placed anywhere in the MR room.
6. Expanded delivery range of the pump should be 0.1 – 1400 ml / hr.
7. The system should be field upgradable on the existing pump.
8. There should be an air inline ultrasonic bubble detector in the upgrade.
9. It should be CE marked and/or FDA approved.
10. It should be certified for use up to a 3T magnet by at-least one major MRI system manufacturer.
11. It should either be manufactured by or approved by Iradimed the original manufacturer of 3860 MR safe pump.

AIHPL/8155/2015-16

Dated: 27/07/2015

**The Head
NMR Department
All India Institute of Medical Sciences
Ansari Nagar
New Delhi- 110029**

Dear Sir

This refers to the recent discussions the undersigned had with your regarding your requirement for adding second channel to your existing 3860 MR safe Pump. As desired our formal offer is detailed below for your kind consideration.

Sl No.	Item Description	Qty	Rate in Indian Rupees	Amount in Indian Rupees
03.	Side car for adding second channel to Iradimed 3860 MR safe pump. (This Side car adds Second Channel to your existing pump for delivery of two drugs simultaneously.)	01 No.	6,40,000-00	6,40,000-00
Total Price not including DVAT				6,40,000-00

TERMS OF OFFER

- Prices :** The prices offered ex – our Delhi stores and are net in Indian Rupees not including ST which shall be charged extra @ 5% (DVAT) or as may be applicable at the time of delivery.
- Payment :** 100% against delivery of the stores.
- Validity :** Our offer is valid for acceptance up to 90 days beyond which it shall be subject to our confirmation.
- Delivery :** We shall arrange for the delivery within 16-20 weeks from the date of receipt of your commercially complete order.
- Guaranty/Warranty :** Five year from the date of supply, against any kind of manufacturing defect.

Thanking you and looking forward to getting your valued patronage.

Yours truly,
for Anaecon India Healthcare Pvt. Ltd.

Vijay Jain
Managing Director



Dated: 27/07/2015

RATE CERTIFICATE

With reference to our offer no. AIHPL/1985/2015-16 date 13/07/2015, against your subject we wish to confirm that we have not supplied the same machine in same denomination, as offered to you, at a rate lower than the quoted rate in the offer to any Govt. organization or any other Institute.

Yours Truly
For Anaecon India Healthcare Pvt. Ltd

Vijay Jain
Managing Director



ANAECON INDIA HEALTHCARE PVT. LTD.
NEW DELHI






सत्यमेव जयते

INDIA NON JUDICIAL

Government of National Capital Territory of Delhi

e-Stamp

Certificate No. : IN-DL23899611194388N
Certificate Issued Date : 04-Jul-2015 12:20 PM
Account Reference : IMPACC (IV)/ dl700703/ DELHI/ DL-DLH
Unique Doc. Reference : SUBIN-DL70070345063784225631N
Purchased by : ANAECON INDIA HEALTHCARE PVT LTD
Description of Document : Article 5 General Agreement
Property Description : Not Applicable
Consideration Price (Rs.) : 0
(Zero)
First Party : ANAECON INDIA HEALTHCARE PVT LTD
Second Party : Not Applicable
Stamp Duty Paid By : ANAECON INDIA HEALTHCARE PVT LTD
Stamp Duty Amount(Rs.) : 10
(Ten only)



.....Please write or type below this line.....

Enclosure to our: AIHPL/8156/2015-16

Dated: 27/07/2015

Non Conviction / Non Black Listing Certificate

I, VIJAY JAIN S/b LATE SH. R.C. JAIN DO SOLEMNLY AFFIRM :-
That I am the Managing Director of M/S Anaecon India P-13, M.I.G. flats Prasad Nagar, New Delhi- 110005.



[Handwritten Signature]

That no Criminal/ BLACK LISTING case is pending against the firm.

That the firm has not been black listed/banned or convicted by any government/Pvt. or quasi-government agency or PSU.



27 JUL 2015


ATTESTED

NOTARY PUBLIC DELHI

PROPRIETARY CERTIFICATE

This is to certify that **MRidium 3860 MR Infusion Pump, its remote and side car (second channel)** are proprietary products of IRadimed Corporation USA and these devices being manufactured by us are compliant for use with all MRI systems from 0.2 to 3.0 Tesla and it can be positioned anywhere in the MR room behind the 10,000 Gauss line.

To the best of our knowledge no other company is making a similar product with the above features.



Todd Weber
Vice President





ADVANCED TECHNOLOGIES IN MR FLUID MANAGEMENT

7457 Aloma Avenue, Suite 201
Winter Park, Florida 32792
Ph: 407.677.8022
Fx: 407.677.5037
www.iradimed.com

August 3rd 2015

TO WHOM IT MAY CONCERN:

IRADIMED CORPORATION hereby confirms having authorised:

Anaecon India Healthcare Pvt. Ltd.
P-13 M.I.G. Flats
Prasad Nagar
New Delhi- 110005
Tel: 011-45051413/14
Fax: 011-45051412

as our exclusive distributor in India for marketing and sales as well as installation and after sales service for our products as per our distributor agreement with them through December 2018.

Sincerely

A handwritten signature in black ink, appearing to read "Todd Weber", is written over a faint, larger signature.

Todd Weber
VP International Sales
IRadimed Corporation
Mobile: +1 770 842 7132
tweber@iradimed.com

