



COMPREHENSIVE RURAL HEALTH SERVICES PROJECT
CENTRE FOR COMMUNITY MEDICINE
ALL INDIA INSTITUTE OF MEDICAL SCIENCES
BALLABGARH - 121004 (HR) / Ph. 0129-2241362/Fax. 0129-2211227

10 JAN 2017

S.R.E.No.15/16-17/CRHSP-AIIMS /Ballabgarh

Dated the: 22.12.2016

For Purchase of : Patient Trolley cum Bed -

The sealed quotations are invited by post/per bearer (under two bid system) in **two separate sealed covers duly marked as "PRICE BID & TECHNICAL BID"** (as the case may be) and both packed in other single envelope mentioning name of the quoted item for the supply of item detailed below which are required for CRHS Project, Ballabgarh.

The quotations duly signed and marked as **"SRE No.15/CRHSP/Ballabgarh/16-17"** (**PRICE BID & TECHNICAL BID separately**) should reach the office of the Professor In-charge, CRHSP, Ballabgarh on or before **96th January, 2017 at 12.00 Noon**. All quotations should be type written or written with indelible ink. All over writing and erased entries will not be considered and treated as deleted entries from the quotations. The rates should be valid at least for one year. **A clause by clause compliance statement should be submitted, failing which the quotation will not be considered.**

The tenders are required to enclose with the quotation an **Earnest Money of Rs.5,000/- (Rupees Five Thousand only)** through a demand draft/pay order drawn in favor of the **"Director, AIIMS"**, failing which the quotation will not be considered. In case of failure to comply with the supply order in terms of the contract, purchase will be made at the cost and risk of the contractors and or liquidated damages may be recovered from them. The period within which the goods can be supplied from the date of issue of the supply order should be specified.

The offers submitted as quality A, B, C, & D etc., will be summarily rejected, any alteration/over writing in the specification / rates shall not be considered at all.

The opening of quotations/tenders (technical bid only) will be intimated to the bidders separately. The quantity shown in the Short Rate Enquiry can be increased or decreased to any extent depending upon the actual requirement.

Terms and Conditions:-

Standard AIIMS terms and conditions apply.

(DR. SHASHI KANT)
PROFESSOR & HEAD,
CCM, CRHS PROJECT, BALLAGARH

The firm should submit the following:-

1. **A clause by clause compliance statement should be submitted.**
2. Price Reasonable Certificate.
3. An undertaking on the official letterhead of the company that the firm has not been black listed.
4. An undertaking on the official letter head of the company that no Vigilance case is pending against the firm.



COMPREHENSIVE RURAL HEALTH SERVICES PROJECT
CENTRE FOR COMMUNITY MEDICINE
ALL INDIA INSTITUTE OF MEDICAL SCIENCES
BALLABGARH - 121004 (HR) / Ph. 0129-2241362/Fax. 0129-2211227

Technical Specifications for Patient Trolley Cum Bed.

1. Should have three sectional mattress base made of X-Ray translucent high pressure laminate with facility to insert X-ray cassette from either sides & ends of the trolley.
2. Should be able to X-Ray the patient from positions along the entire length and width of the trolley.
3. Should have pneumatic step less adjustment for back section, Trendelenburg, Reverse Trendelenburg and foot section.
4. Should have hydraulic height adjustment with a foot paddle on either side of the trolley.
5. Frame of the trolley should move with mattress base when foot section/back section is adjusted.
6. Frame should be made up of Epoxy Powder Coated Steel.
7. Should have central breaking system with steering facility.
8. Should be Equipped with heavy duty castors diameter 200mm.
9. Should have bumper at all the four corners of the trolley.
10. Should have facility to fix IV Rod at all the four corners and middle of mattress base frame.
11. Should have place for fixing 'B' Type Oxygen Cylinder.

DIMENSIONS

12. * **LENGTH** :- 2120 mm
- * **WIDTH** :- 800 mm
- * **HEIGHT** :- 550-930 mm
13. Trendelenburg :- 10-20 Step less.

Total Quantity required = 02

14. Anti Trendelenburg :- 7-10 Step Less.
15. Weight capacity :- up to 200 Kgs.
16. X-Ray viewing area :- Entire Length.

SHOULD SUPPLY WITH THE FOLLOWING ACCESSORIES

17. Anti Static Hygiene mattress (80 mm) with pull straps, 01 pc.
18. Collapsible side rails, 01 pc.
19. IV Rod :- 02 pcs.
20. Cylinder holder for B-type oxygen cylinder :- 01 pc.

Comprehensive Warranty for 05 (Five Years) date of instalatioo.

"The Vendor should ensure preventive maintenance biannually at the site of commission. A penalty of Rs.1000/- per day would be imposed on violation of this clause. "