

DEPARTMENT OF PULMONARY, CRITICAL CARE & SLEEP MEDICINE

AIIMS, NEW DELHI – 110029

- (i) Sub: Purchase of "Nocolyse Disinfectant Equipment" 02 nos. For the Department of PCC&SM at AIIMS, New Delhi-110029 on proprietary basis inviting comments thereon.

T. No. 08/PCCSM/AM/2020-21

- (ii) The institute is in the process to purchase from "M/s S. J. Mega Solutions, New Delhi. The PAC certifications by M/s S. J. Mega Solutions, New Delhi as well as the user department are attached.

The above documents are being uploaded for open information to submit objection, comments, if any, from any, manufacture regarding proprietary nature of the equipment/item within 15 day from the date of issue/uploading of the notification. The comments should be received in office of the Head, Deptt. Of PCC&SM, AIIMS New Delhi-110029 on or before ~~11.09.2020~~ upto 4:30PM failing which it will be presumed that any other vendor is having documents to offer and case will be decided on merits.

Yours faithfully



Head

Department of PCC&SM


Encl. Related Documents


**ALL INDIA INSTITUTE OF MEDICAL SCIENCES
ANSARI NAGAR, NEW DELHI-110029**

PROPRIETORY/ SPECIFIC BRAND GOODS CERTIFICATE

- | | | |
|----|---|---|
| 1. | Item/Type/Model No. required | Nocolse Disinfectant Equipment |
| 2. | Is the item a spare parts attachment or accessory for an existing equipment. | Yes |
| 3. | Name of the manufacturers/supplier of the item proposed by the Indentor. | M/s S. J. Mega Solutions
B-1341/42, GD Colony mayar Vihar Phase-III,
New Delhi - 110096 |
| 4. | Are they sole manufacturers/sold distributors of the item. | Yes (as the certificate provided by the fir |
| 5. | Is there any other item with similar/ specification available in the market to meet the job requirement envisaged, if the answer is yes, why the same can't be procured. Demanding officer should bring out comparative functional advantages/cost effectiveness of the recommended item from these offered by other. | No (Nocolse Disinfectant Equipment)

requirements of the specifications
of the item needs for the Department of PMSD |
| 6. | What were the efforts made to locate alternative source of supply or use other substitutes. | --- |
| 7. | Why open/limited tender can't be resorted to, for locating alternative source. | As the firm is sole Proprietor |
| 8. | Are the proprietary items certifying that the rates are reasonable or not. | Yes the rates are reasonable |
| 9. | Any other justification for procuring item from single source. | No |


**Signature of Indentor
(Demanding Officer)**


**COUNTERSIGNE
(Head of the Department)**

I certify that the item at Sr. No. 1 above is required to be procured on single tender basis as the supply definitely known/ the specified brand proposed was advantages in meeting our functional requirements and limited tender system could be dispensed with as they would serve no useful purpose in this particular case.

(Strike out whichever is not applicable)

Department of Pulmonary, Critical Care & Sleep Medicine
AIIMS, New Delhi

Qty - 02 No.

Nocosprey -2 – Nocolyse Disinfectant Equipment

1. Nocosprey has a heating and Ionizing Turbine which diffuses H₂O₂ base disinfectant in OH ions thus eliminates water completely.
2. Adheres to the highest and latest disinfectant norms – NFT 72881 (2014)
3. CE marked equipment manufactured as per ISO 13485 standards.
4. Diffusion at rate of 80 meter per second at 40° Celsius
5. Weight of Nocosprey-2 is approx 5.5 to 6.5 kg easily moveable
6. 1,100 watt electrical turbine rotates on 2,200 RPM
7. 100% Bio-degradable, Non Corrosive, Resistance to discoloration of installed equipment
8. Completely eliminates Bacteria, fungus, virus and spores
9. Delay start and automatic stop equipped machine
10. High speed ensures uniform sedimentation on each square cm of treated area.
11. 1 ml for 1 cubic meter and can treat volume from 10 to 1000 cubic meters.
12. Only 30 minutes contact time to reach optimum level of efficacy.
13. No residue is left after the treatment. Water as base for contamination is completely eliminated.
14. Germs/pathogens do not create resistance to Nocolyse.
15. Only 3-4 minutes of diffusion time required to treat an OT of 70 cubic meters.

WARRANTY

1. As per AIIMS rules and should be quoted with 2-year comprehensive warranty (including all spares, batteries, circuit and other accessories) and another five 03 years of maintenance warranty.

Sanjiv *Mohan* *✓* *Sanjay* *Pawan*

Dr. Vijay Hadda
MD, PhD., MAMS, FRCP (LONDON), FCCP (USA) — MD (Medicine), FCCP (USA), FISCMM
Professor & Head
Deptt. of Pulmonary, Critical Care & Sleep Medicine
Associate Professor
All India Institute of Medical Sciences
Delhi - 110029

[Signature]

Contd. Pg. 2

Dr. Jitender Sodhi
MD, PhD., MAMS, FRCP (LONDON), FCCP (USA) — MD (Medicine), FCCP (USA), FISCMM
Professor
Deptt. of Pulmonary, Critical Care & Sleep Medicine
Associate Professor
All India Institute of Medical Sciences
Delhi - 110029

::2::

2. The cost of additional accessories should quoted in the price bid which will be used to calculate L1.
3. Prices of all the spare/accessories required for maintenance of the equipment should be quoted along with the price bid. If the firm fails to provide price of any part/accessories in the price bid the same will be provide free of cost by the company if required.
4. Company will be responsible to maintain equipment and accessories in working condition irrespective of the cause/reasons/conditions/nature for not working of the equipment/accessories. Company will replace all the required parts/spares /accessories, labour/service during WARRANTY/CAMC period without any extra charges/taxes" (including and not limited to physical damage).
5. In case of any breakdown, fault, repair should be undertaken within 48 hours of receipt of such information. Failure to do so shall make the company liable for a penalty of Rs.1000/- per day.

Dr. ANANT MOHAN
MD, PhD., MAMS, FRCP (LONDON), FCCP (USA)
Professor & Head
Deptt. of Pulmonary, Critical Care & Sleep Medicine
All India Institute of Medical Sciences
New Delhi - 110029

Jawahar

[Signature]

[Signature]

Pawan

[Signature]

Atinder Sodhi
Assistant Professor
Hospital Administration
All India Institute of Medical Sciences
New Delhi - 110029

अंतरा
अतिरिक्त
All India Institute of Medical Sciences



829 Rue Marcel Paul | 94500 Champigny-sur-Marne | France
+ 33 (0)1 45 18 78 70 commercial@oxypharm.net

PROPRIETARY CERTIFICATE

TO WHOM SO EVER IT MAY CONCERN

We, OXY'PHARM, FRANCE with office 829, rue Marcel Paul - ZA des Grands Godets, 94500 Champigny-sur-Marne is the sole worldwide manufacturer of the NOCOSPRAY-NOCOLYSE DISINFECTION EQUIPMENT.

No company other than OXY'PHARM, FRANCE manufactures NOCOSPRAY-NOCOLYSE DISINFECTION EQUIPMENT of similar configuration.

Further, NOCOSPRAY is patented equipment with patent number 0700964 (Publication Number 2912328) on the name of Dr. Thierry Rouleau, who is the shareholder in OXY'PHARM, FRANCE.

OXY'PHARM
829, Rue Marcel Paul
94500 Champigny-sur-Marne
T. : 01 45 18 78 70
M. : commercial@oxypharm.net
RCS Creteil : 448 974 253

A handwritten signature in black ink, consisting of a series of loops and strokes, positioned below the company contact information.

A second handwritten signature in black ink, located at the bottom right of the page.



829 Rue Marcel Paul • 94500 Champigny-sur-Marne • France
+33 (0)1 45 18 78 70 @ commercial@oxypharm.net

TO WHOMSOEVER IT MAY CONCERN
MANUFACTURER'S AUTHORIZATION

We, M/s. OXY'PHARM, France having our office/factory at 829 rue Marcel Paul - 94500 Champigny sur-Marne, Paris, France are the sole manufacturer of proprietary product – Nocospray, heating and ionising turbine based disinfection system.

We hereby authorize M/s. **SJ Mega Solutions, B-1341/42, GD Colony, Mayur Vihar, Phase 111, New Delhi – 110096** to bid, negotiate and conclude the contract for the items manufactured by us in All India Institute of Medical Sciences, New Delhi.

OXY'PHARM
829, Rue Marcel Paul
94500 Champigny-sur-Marne
T. : 01 45 18 78 70
M. : commercial@oxypharm.net
RCS Creteil : 448 974 253

SARL OXYPHARM
Capital de 300 000 Euros
SIREN 448 974 253
RCS Creteil
APE 4646Z
N° TVA FR 74 448 974 253

www.oxypharm.net