

Laminar Hood specifications:

1. Should have ~80% air circulation and balance exhausted.
2. Should have front door of clear Perspex in two sections, with hinges. Sloped Front Cabinet front is sloped 10° for enhanced comfort and reduced operator fatigue.
3. Should have table top single piece work tray of stainless steel of AISI- 304 quality along with stainless steel stand with lockable wheel
4. Should have at least exterior Dimensions HxWxD inches 55 x 50 x 30 and Interior Dimensions HxWxD inches (mm) 28 x 45 x 15
5. Should have U.V germicidal tube fitted in the work area and in the exhaust plenum before the exhaust HEPA filter.
6. Reduced Noise Level Lower noise level (should have sound Pressure Level not more than 65 dB) enhances the attention and allows user to focus on work without distraction.
7. Displays real-time information of overall cabinet status including operation safety and service requirements
8. Visual and audible alarm indicates whether front window is in correct working position
9. Clearly displays critical safety and real-time performance data including inflow and down flow velocities and hours of operation
10. Audible and visual alarms indicate when airflow is safe or restricted
11. Displays reduced speed operation when front window is closed
12. Easy Servicing Fan control and power supply can be replaced independently of the DC motor with no need for disruptive decontamination of the cabinet. All cabinet components, including HEPA filters, are easily accessible from the front to allow for rapid service and minimal work disruption.
13. Static pressure inclined tube Manometer.
14. HEPA filters (0.3 micron) on the exhaust side.

डॉ. टोनी जॉर्ज जेकब
Dr. TONY GEORGE JACOB, MD, Ph.D.
सहायक आचार्य/Assistant Professor
शरीर रचना विभाग/Deptt. of Anatomy
अ.भा.आ.सं., नई दिल्ली-29/A.I.I.M.S.

Subhas

T. B. S.

R. Singh
Professor
Department of Anatomy
M.S. New Delhi-29

Dr. A. Shariff
आचार्य/Professor
शरीर रचना विभाग/Deptt. of Anatomy
अ.भा.आ.सं., नई दिल्ली/A.I.I.M.S., New Delhi-29

Dr. Punit Ka

डॉ. अमित लठवाल/Dr. Amit Lathwal
सहायक आचार्य/Assistant Professor
प्रशासन विभाग/Dept. of Hospital Administration
अखिल भारतीय आयुर्विज्ञान संस्थान/AIIMS
अंसारी नगर, नई दिल्ली/Ansari Nagar, N.D.-110029

डॉ. टी. एस. रॉय/Dr. T. S. Roy
आचार्य एवं अध्यक्ष/Professor & Head
शरीर रचना विभाग/Dept. of Anatomy
अ.भा.आ.सं., नई दिल्ली, भारत/AIIMS, New Delhi, India

डॉ. पुनीत कौर / Dr. Punit Ka
आचार्य एवं अध्यक्ष / Professor &
जैवभौतिकी विभाग / Dept. of Bio
अ.भा.आ.सं. / AIIMS

15. All technical claims should be supported by product catalogue, public website of the manufacturer.
16. Should be completely complying the point no 18.1 and 18.6 at page no. 16 of the AIIMS tender documents.
17. The system should be certified from European CE / UL / ISO 9001 for quality assurance.
18. The firm should have service center in Delhi and well-trained engineers for technical support in case of a breakdown.
19. Standard instrument comprehensive warranty for 5 years + 5 years comprehensive maintenance contract (CMC) after completion of 5 years warranty.
20. Firm should provide the list of spares and accessories, which are included in comprehensive warranty and CMC.
21. Firm should also provide the list of consumables, which are not included in the comprehensive warranty and CMC. Company should quote the prices (freeze for next five years) of each consumable separately, which are not included in CMC.
22. Provide point-wise compliance statement indicating compliance of the technical specification vis a vis technical specification offered. Please indicate page number/location of the compliance in the technical brochure.
23. Equipment and power supply connector should be compatible with Indian power supply system
24. Performance Certificate from at least 3 users preferably from Government sector, should be provided.

Subhash

Bar

Shariff



डॉ. सुभाष चन्द्र यादव
Dr. SUBHASH CHANDRA YADAV
सहायक आचार्य/Assistant Professor
शरीर रचना विभाग/Department of Anatomy
अ.भा.आ.सं., नई दिल्ली-29/A.I.I.M.S., New Delhi-29

डॉ. ए. शरीफ/Dr. A. SHARIFF
आचार्य/Professor
शरीर रचना विभाग/Deptt. of Anatomy
अ.भा.आ.सं., नई दिल्ली/AIIMS, New Delhi-29

डॉ. टोनी जॉर्ज जेकर
Dr. TONY GEORGE JACOB MD, PhD
सहायक आचार्य/Assistant Professor
शरीर रचना विभाग/Deptt. of Anatomy
अ.भा.आ.सं., नई दिल्ली-29/A.I.I.M.S., New Delhi-29

hantian

डॉ. पुनीत कौर / Dr. Punit Kaur
आचार्य एवं अध्यक्ष / Professor & Head
जैवभौतिकी विभाग / Deptt. of Biophysics
अ.भा.आ.सं. / A.I.I.M.S.
अंसारी नगर, नई दिल्ली-29 / Ansari Nagar, New Delhi-29

रघुचंद्रांगिरा
Dr. Raghunath Singh
Professor
Department of Anatomy
A.I.I.M.S., New Delhi-29

डॉ. अमित लठवाल/Dr. Amit Lathwal
सहायक आचार्य/Assistant Professor
प्रशासन विभाग/Dept. of Hospital Administration
अखिल भारतीय आयुर्विज्ञान संस्थान/AIIMS
अंसारी नगर, नई दिल्ली/Ansari Nagar, N.D.-110029



डॉ. टी. एस. रॉय/Dr. T. S. ROY, MD PhD
आचार्य एवं अध्यक्ष/Professor & Head
शरीर रचना विभाग/Dept. of Anatomy
अ.भा.आ.सं., नई दिल्ली, भारत/AIIMS, New Delhi, India

Sharma

Terms & Conditions:

- The quotations should be submitted to the Professor & Head, Department of Anatomy, Room No. 1010, AIIMS, New Delhi-29 and sent by post/per bearer by 7th October, 2016 up to 12.00 P.M. Quotation should be sealed in an envelope and the "name of the item" clearly written on top of the envelop.
- The make of the article offered should invariably be quoted. Quotation should be typed written in ink. All overwriting and erased entries will be deleted from the quotation. The rates should be valid for at least THREE MONTHS.
- VAT/Sales tax or any other kind of tax(s) must be separately mentioned against each item. In case no sales tax is chargeable, prices must be quoted as NET.
- **Warranty/Guarantee:** two years followed by three years CMC comprehensive warranty.
- **Installation:** The firm install the equipment on the site, free of cost
- The firm should be able to give demonstration of the item if required within 5 days of requisition.
- The firm must supply the item within 20 days after confirmation of order.
- **The successful bidder has to submit performances bank guarantee of 10% of the order value valid upto warranty period.**
- The payment will be made electronically viz RTGS/NEFT against delivery and satisfactory installation of the item. The following information should be also mentioned in the quotation.
- Name of the beneficiary:
- Account No. of the beneficiary:
- IFCS code of the bank/branch:
- The quantity can be increased or decreased to any extent depending upon the actual requirement.
- The Professor & Head of the Department reserves the right to cancel/reject full or any part of the rate enquiry which generally do not fulfil the conditions stipulated in the rate enquiry.
- Payment will be made after receiving the items in good condition and satisfactory installation.
- No extra transportation charge will be given to the firm.
- The firm must supply the items within the date mentioned on the supply order copy.

Thanking you

Renu Dhingra

(Dr. Renu Dhingra)

Professor

Computer facility: With request for uploading the same in AIIMS websites for the period of 15 Days.

Dr. Renu Dhingra
Professor
Department of Anatomy
A.I.I.M.S., New Delhi-29