


## Rocking Shaker

1. It should have rocking motion
2. It should have steady motion at 5rpm with a Maximum tilt angle 12-15°
3. It should be controlled by a Digital microprocessor and digital LC display
4. Speed range should be from 5 to 100rpm with increments of 1rpm
5. The Timer should be able to clock from 1 to 9999 (minutes) with alarm for completion of programme. It should also be able to run on Continuous mode.
6. It should have a maximum load capacity of 8kg
7. Motor should be a DC brushless type. With total noise level less than 10dB.
8. It should be able to operate at temperatures from Ambient to 40°C
9. Platform dimension should be not less than (WxL) 30x30cm
10. Platform material should be made of powder coated steel, with clamps and corrugated rubber sheet to prevent slippage of boxes and plates
11. It should have a Stacking platform
12. It should not weigh more than 7 Kg
13. It should be made of a Flame retardant material and painted iron metal bottom plates
14. Rated voltage should be 110V / 220V (selectable)
15. Features in the quotations should be substantiated with proper Principal Company Catalogue.
16. Should operate under a Quality Management System which complies with the requirements of ISO 9001:2008 for design, manufacture and services.
17. Copies of all certifications e.g. Quality Standard certificate, Propriety Item/parts, Patent of parts/ technology, Principal company/Authorised Distributorship should be attached with the quotations.
18. In-House Service Engineers from principal company/ or authorised agents should be available on one-day notice basis in case of emergency.
19. Only Principal Companies or Authorized Distributors from Principal companies should quote. Quotations from non - authorized distributors will not be entertained.
20. Compliance sheet should be attached along with the quotation. Compliant points should be highlighted in the company catalogue. Compliant points should be given (in the sheet) in order of the specifications' serial order

*2008-2009 Reg*

  
डॉ० टोनी जॉर्ज जेकब  
Dr. TONY GEORGE JACOB, MD, Ph.D.  
सहायक आचार्य/Assistant Professor  
शरीर रचना विभाग/Deptt. of Anatomy  
अ.भा.आ.सं., नई दिल्ली-29/A.I.I.M.S., New Delhi-29

  
डॉ० ए. शरीफ/Dr. A. SHARIFF  
आचार्य/Professor  
शरीर रचना विभाग/Deptt. of Anatomy  
अ.भा.आ.सं., नई दिल्ली/AIIMS, New Delhi-29

  
Dr. Renu Dhingra  
Professor  
Department of Anatomy  
नई दिल्ली-29



डॉ० टोनी जॉर्ज जेकब  
Dr. TONY GEORGE JACOB, MD, Ph.D.  
सहायक आचार्य/Assistant Professor  
शरीर रचना विभाग/Deptt. of Anatomy  
अ.भा.आ.सं., नई दिल्ली-29/A.I.I.M.S., New Delhi-29

### Terms & Conditions:

- The quotations should be submitted to the Professor & Head, Department of Anatomy, Room No. 1010, AIIMS, New Delhi-29 and sent by post/per bearer by 7<sup>th</sup> October, 2016 up to 12.00 P.M. Quotation should be sealed in an envelope and the "name of the item" clearly written on top of the envelop.
  - The make of the article offered should invariably be quoted. Quotation should be typed written in ink. All overwriting and erased entries will be deleted from the quotation. The rates should be valid for at least THREE MONTHS.
  - VAT/Sales tax or any other kind of tax(s) must be separately mentioned against each item. In case no sales tax is chargeable, prices must be quoted as NET.
  - **Warranty/Guarantee:** two years followed by three years CMC comprehensive warranty.
  - **Installation:** The firm install the equipment on the site, free of cost
  - The firm should be able to give demonstration of the item if required within 5 days of requisition.
  - The firm must supply the item within 20 days after confirmation of order.
  - **The successful bidder has to submit performances bank guarantee of 10% of the order value valid upto warranty period.**
  - The payment will be made electronically viz RTGS/NEFT against delivery and satisfactory installation of the item. The following information should be also mentioned in the quotation.
  - Name of the beneficiary:
  - Account No. of the beneficiary:
  - IFCS code of the bank/branch:
  - The quantity can be increased or decreased to any extent depending upon the actual requirement.
  - The Professor & Head of the Department reserves the right to cancel/reject full or any part of the rate enquiry which generally do not fulfil the conditions stipulated in the rate enquiry.
  - Payment will be made after receiving the items in good condition and satisfactory installation.
  - No extra transportation charge will be given to the firm.
  - The firm must supply the items within the date mentioned on the supply order copy.
- Thanking you

*Renu Dhingra*  
(Dr. Renu Dhingra)

Professor

Computer facility: With request for uploading the same in AIIMS websites for the period of 15 Days.

Dr. Renu Dhingra  
Professor  
Department of Anatomy  
A.I.I.M.S., New Delhi-29