

Specification for Water Bath:
It should have/be

1. Temperature setting range °C: Ambient +5°C to 99°C in 0.1 steps
2. Volume range: 12 liters
3. Temperature Stability (DIN 12876*) @ 37°C: ±0.2
4. Digital display and 3 digit wide angle view temperature display.
5. Stainless steel tank high grade steel with durable polished finish.
6. Drain tap included within the system.
7. Raised feet which allow lifting whilst holding base of tank.
8. Three programmable temperature presets – should permit commonly used temperatures to be quickly selected.
9. Provided with front panel lock-out which disables front panel controls and prevents settings being accidentally changed.
10. Countdown timer with audible buzzer for accurate reaction timing.
11. Fixed thermal cut-out which protects the bath to accidentally run without water.
12. User settable, over-temperature sample protection features for protecting samples from overheating.
13. Digital PID control for quick heat-up and precision control throughout the temperature range.
14. Heater mat sensor bonded to underside of tank for optimizing temperature uniformity and workspace.
15. Non-drip polycarbonate lid included as standard to improve performance, limits evaporation and conserves energy.
16. Chemical and abrasion resistance polycarbonate perforated base tray included as standard.
17. Single or dual-point user calibration for optimum re-calibration of working temperature.
18. CE/EC(European conformity) USFDA certified
19. Warranty AMC/CMC according to AIIMS rules

Renu Dhiria
Dr Renu Dhiria
Professor
Department of Anatomy
AIIMS, New Delhi-29

A. Shariff
डॉ० ए. शरीफ़ / Dr. A. SHARIFF
आचार्य / Professor
शरीर रचना विभाग / Deptt. of Anatomy
अ.मो.आ.सं., नई दिल्ली / AIIMS, New Delhi-29

Dr. T. George Jacob
डॉ० टोनी जॉर्ज जेकब
TONY GEORGE JACOB, MD, Ph.D.
सहायक आचार्य / Assistant Professor
शरीर रचना विभाग / Deptt. of Anatomy
अ.मो.आ.सं., नई दिल्ली / AIIMS, New Delhi-29

Terms & Conditions:

- The quotations should be submitted to the Professor & Head, Department of Anatomy, Room No. 1010, AIIMS, New Delhi-29 and sent by post/per bearer by 7th October, 2016 up to 12.00 P.M. Quotation should be sealed in an envelope and the "name of the item" clearly written on top of the envelop.
- The make of the article offered should invariably be quoted. Quotation should be typed written in ink. All overwriting and erased entries will be deleted from the quotation. The rates should be valid for at least THREE MONTHS.
- VAT/Sales tax or any other kind of tax(s) must be separately mentioned against each item. In case no sales tax is chargeable, prices must be quoted as NET.
- **Warranty/Guarantee:** two years followed by three years CMC comprehensive warranty.
- **Installation:** The firm install the equipment on the site, free of cost
- The firm should be able to give demonstration of the item if required within 5 days of requisition.
- The firm must supply the item within 20 days after confirmation of order.
- **The successful bidder has to submit performances bank guarantee of 10% of the order value valid upto warranty period.**
- The payment will be made electronically viz RTGS/NEFT against delivery and satisfactory installation of the item. The following information should be also mentioned in the quotation.
- Name of the beneficiary:
- Account No. of the beneficiary:
- IFCS code of the bank/branch:
- The quantity can be increased or decreased to any extent depending upon the actual requirement.
- The Professor & Head of the Department reserves the right to cancel/reject full or any part of the rate enquiry which generally do not fulfil the conditions stipulated in the rate enquiry.
- Payment will be made after receiving the items in good condition and satisfactory installation.
- No extra transportation charge will be given to the firm.
- The firm must supply the items within the date mentioned on the supply order copy.

Thanking you

Renu Dhingra

(Dr. Renu Dhingra)

Professor

Computer facility: With request for uploading the same in AIIMS websites for the period of 15 Days.

Dr. Renu Dhingra
Professor
Department of Anatomy
A.I.I.M.S., New Delhi-29